

West End Grange
Epworth Water Gardens
West End Road
Epworth
Doncaster
DN9 1LA



Certificate No:

51580

Tel: 01427 875 867
Web: www.allweighs.co.uk

Calibration Certificate of Accuracy

Business Name:

Ward Brothers. (Steel) Ltd

Address:

Cleveland House
Cleveland Street
Darlington

Postcode:

DL1 2NU

Contact Name:

Mr Billy Ward

Test Date:

21 August 2015

Location:

In Weighbridge

Manufacturer:

Revere VT200

Serial/Plant No:

T13110246

Capacity:

50000

kg

Increment:

20

kg

Frequency:

12

months

Next Due:

21 August 2016

As Found

Measurement Points Reading Deviation

0	0	0
8000	8000	0
16000	16000	0
20840	20840	0
33540	33540	0
45620	45620	0

Repeatability

45620	45620	0
-------	-------	---

Off-centre Loading

6840	6840	0
------	------	---

Engineer Name:

J Scott

Definitive

Measurement Points Reading Deviation

0	0	0
8000	8000	0
16000	16000	0
20840	20840	0
33540	33540	0
45620	45620	0

Repeatability

45620	45620	0
-------	-------	---

Off-centre Loading

6840	6840	0
------	------	---

Trading Standards Maximum Tolerance

60

+/-

60

kg

Traceable Weights Used:

W14035



West End Grange
Epworth Water Gardens
West End Road
Epworth
Doncaster
DN9 1LA



Certificate No:

51581

Tel: 01427 875 867
Web: www.allweighs.co.uk

Calibration Certificate of Accuracy

Business Name:

Ward Brothers. (Steel) Ltd

Address:

Cleveland House
Cleveland Street
Darlington

Postcode:

DL1 2NU

Contact Name:

Mr Billy Ward

Test Date:

21 August 2015

Location:

Out Weighbridge

Manufacturer:

Revere VT200

Serial/Plant No:

T13110243

Capacity:

50000

kg

Increment:

20

kg

Frequency:

12

months

Next Due:

21 August 2016

As Found

Measurement Points Reading Deviation

0	0	0
8000	8000	0
16000	16000	0
20840	20840	0
33540	33540	0
45640	45640	20

Repeatability

45640	45640	0
-------	-------	---

Off-centre Loading

6840	6840	0
------	------	---

Engineer Name:

J Scott

Definitive

Measurement Points Reading Deviation

0	0	0
8000	8000	0
16000	16000	0
20840	20840	0
33540	33540	0
45640	45640	0

Repeatability

45640	45640	0
-------	-------	---

Off-centre Loading

6840	6840	0
------	------	---

Trading Standards Maximum Tolerance

60


+/-

60

kg

Traceable Weights Used:

W14035



West End Grange
Epworth Water Gardens
West End Road
Epworth
Doncaster
DN9 1LA



Certificate No:

51582

Tel: 01427 875 867
Web: www.allweighs.co.uk

Calibration Certificate of Accuracy

Business Name:

Ward Brothers. (Steel) Ltd

Address:

Cleveland House
Cleveland Street
Darlington

Postcode:

DL1 2NU

Contact Name:

Mr Billy Ward

Test Date:

21 August 2015

Location:

Non Ferrous

Manufacturer:

LEON LD5204

Serial/Plant No:

100283476

Capacity:

3000

kg

Increment:

1

kg

Frequency:

12

months

Next Due:

21 August 2016

As Found

Measurement Points	Reading	Deviation
--------------------	---------	-----------

0	0	0
1000	1000	0
2000	2000	0
3000	2999	1

Repeatability

2000	2000	0
------	------	---

Off-centre Loading

500	500	0
-----	-----	---

Engineer Name:

J Scott

Definitive

Measurement Points	Reading	Deviation
--------------------	---------	-----------

0	0	0
1000	1000	0
2000	2000	0
3000	3000	0

Repeatability

2000	2000	0
------	------	---

Off-centre Loading

500	500	0
-----	-----	---

Trading Standards Maximum Tolerance

3


+/-

3

kg

Traceable Weights Used:

W14035



West End Grange
Epworth Water Gardens
West End Road
Epworth
Doncaster
DN9 1LA



Certificate No:

51583

Tel: 01427 875 867
Web: www.allweighs.co.uk

Calibration Certificate of Accuracy

Business Name:

Ward Brothers. (Steel) Ltd

Address:

Cleveland House
Cleveland Street
Darlington

Postcode:

DL1 2NU

Contact Name:

Mr Billy Ward

Test Date:

21 August 2015

Location:

Non Ferrous

Manufacturer:

LEON LD5204

Serial/Plant No:

100406391

Capacity:

3000

kg

Increment:

1

kg

Frequency:

12

months

Next Due:

21 August 2016

As Found

Measurement Points	Reading	Deviation
--------------------	---------	-----------

0	0	0
1000	1000	0
2000	2000	0
3000	3000	0

Repeatability

2000	2000	0
------	------	---

Off-centre Loading

500	500	0
-----	-----	---

Engineer Name:

J Scott

Definitive

Measurement Points	Reading	Deviation
--------------------	---------	-----------

0	0	0
1000	1000	0
2000	2000	0
3000	3000	0

Repeatability

2000	2000	0
------	------	---

Off-centre Loading

500	500	0
-----	-----	---

Trading Standards Maximum Tolerance

3

+/-

3

kg

Traceable Weights Used:

W14035