

West End Grange
Epworth Water Gardens
West End Road
Epworth
Doncaster
DN9 1LA



Certificate No:

51577

Tel: 01427 875 867
Web: www.allweighs.co.uk

Calibration Certificate of Accuracy

Test Date: 21 August 2015

Business Name: Ward Brothers. (Steel) Ltd

Address: Quality Row
Byker
Newcastle Upon Tyne
Tyne And Wear

Postcode: NE6 1LB

Contact Name: Mr Nathan Ward

Location: Out Weighbridge

Manufacturer: Avery E1205

Serial/Plant No: 084150256

Capacity: 50000 kg

Increment: 20 kg

Frequency: 12 months

Next Due: 21 August 2016

As Found

Measurement Points Reading Deviation

0	0	0
8000	8000	0
16000	16000	0
20840	20840	0
33560	33560	0
45640	45640	0

Definitive

Measurement Points Reading Deviation

0	0	0
8000	8000	0
16000	16000	0
20840	20840	0
33560	33560	0
45640	45640	0

Repeatability

45640	45640	0
-------	-------	---

Repeatability

45640	45640	0
-------	-------	---

Off-centre Loading

6840	6840	20
------	------	----

Off-centre Loading

6840	6840	0
------	------	---

Engineer Name: J Scott

Trading Standards Maximum Tolerance

60


+/-

60

kg

Traceable Weights Used:

W14035



West End Grange
Epworth Water Gardens
West End Road
Epworth
Doncaster
DN9 1LA



Certificate No:

51578

Tel: 01427 875 867
Web: www.allweighs.co.uk

Calibration Certificate of Accuracy

Test Date: 21 August 2015

Business Name: Ward Brothers. (Steel) Ltd

Location: Non Ferrous

Address: Quality Row
Byker
Newcastle Upon Tyne
Tyne And Wear

Manufacturer: VISHAY VT200

Serial/Plant No: T10110328

Capacity: 3000 kg

Increment: 1 kg

Frequency: 12 months

Next Due: 21 August 2016

Postcode: NE6 1LB

Contact Name: Mr Nathan Ward

As Found

Measurement Points Reading Deviation

0	0	0
500	500	0
1000	1000	0
1500	1500	0
2000	2000	0
3000	3000	0

Definitive

Measurement Points Reading Deviation

0	0	0
500	500	0
1000	1000	0
1500	1500	0
2000	2000	0
3000	3000	0

Repeatability

2000	2000	0
------	------	---

Repeatability

2000	2000	0
------	------	---

Off-centre Loading

500	500	0
-----	-----	---

Off-centre Loading

500	500	0
-----	-----	---

Engineer Name: J Scott


Trading Standards Maximum Tolerance

3 +/-

3 kg

Traceable Weights Used:

W14035



West End Grange
Epworth Water Gardens
West End Road
Epworth
Doncaster
DN9 1LA



Certificate No:

51579

Tel: 01427 875 867
Web: www.allweighs.co.uk

Calibration Certificate of Accuracy

Test Date: 21 August 2015

Business Name: Ward Brothers. (Steel) Ltd

Address: Quality Row
Byker
Newcastle Upon Tyne
Tyne And Wear

Postcode: NE6 1LB

Contact Name: Mr Nathan Ward

Location: In Weighbridge

Manufacturer: Avery L225

Serial/Plant No: 69042007

Capacity: 50000 kg

Increment: 20 kg

Frequency: 12 months

Next Due: 21 August 2016

As Found

Measurement Points Reading Deviation

0	0	0
8000	8000	0
16000	16000	0
20840	20840	0
33540	33540	0
45640	45660	20

Repeatability

45640	45660	20
-------	-------	----

Off-centre Loading

6840	6840	0
------	------	---

Engineer Name: J Scott

Definitive

Measurement Points Reading Deviation

0	0	0
8000	8000	0
16000	16000	0
20840	20840	0
33540	33540	0
45640	45640	0

Repeatability

45640	45640	0
-------	-------	---

Off-centre Loading

6840	6840	0
------	------	---

Trading Standards Maximum Tolerance

60

+/-

60

kg

Traceable Weights Used:

W14035