

West End Grange **Epworth Water Gardens** Epworth Doncaster



20834

West End Road DN9 1LA

Calibration Certificate of Accuracy Tel: 01427 875 867 Fax: 01427 728 035 Web: www.allweighs.co.uk 21 August 2015 Test Date: Sunderland weighbridge Location: **Business Name:** Ward Brothers Sunderland Address: Robinson Terrace Manufacturer: Vishay VT300 Sunderland Serial/Plant No: T12310159 50000 kg Capacity: 20 Increment: kg months Postcode: SR2 8GW 12 Frequency: Contact Name: Larry Ward Next Due: 21 August 2016 As Found **Definitive** Measurement Points Measurement Points Reading Deviation Reading Deviation 0 0 lo 0 0 0 8000 8000 0 8000 8000 0 16000 16000 0 16000 16000 0 0 20840 20840 0 20840 20840 33560 0 33560 33560 33560 0 Ю 0 45660 45660 45660 45660 Repeatability Repeatability 45660 llо 45660 0 45660 45660 Off-centre Loading Off-centre Loading

Engineer Name: Kevin Scott

12840

12840

0

Trading Standards Maximum Tolerance

0

30 +/kg

Traceable Weights Used:

12840

12840

6550

QMF17 Issue 2 23.02.2013



West End Grange Epworth Water Gardens West End Road Epworth Doncaster DN9 1LA



Test Date:

20835

Tel: 01427 875 867 Fax: 01427 728 035 Web: www.allweighs.co.uk

Calibration Certificate of Accuracy

21 August 2015 Location: Sunderland new weighbridge Ward Brothers Sunderland **Business Name:** Manufacturer: Address: Robinson Terrace Dini Argeo Sunderland Serial/Plant No: 3590 50000 kg Capacity: 20 Increment: kg months Postcode: SR2 8GW Frequency: 12 Contact Name: Larry Ward Next Due: 21 August 2016 **Definitive**

As Found

Measurement Points	Reading	Deviation	Measurement Point	s Reading	Deviation	
0	0	0	0	0	0	
8000	8000	0	8000	8000	0	
16000	16000	0	16000	16000	0	
20840	20840	0	20840	20840	0	
33560	33560	0	33560	33560	0	
45660	45660	0	45660	45660	0	
Repeatability			Repeatability			
45660	45660	0	45660	45660	0	
Off-centre Loading			Off-centre Loading			
12840	12840	0	12840	12840	0	
Engineer Name:	Kevin Scott			Trading Standards Maximum Tolerance		



Traceable Weights Used:

30 kg 6550